



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 8 2005 INITIALS: JEC OWNER: WSP
DISTRICT: 3 ADDRESS: T.I. 82N MP 11
LOCATION: Plymouth POE #54 COUNTY: Benton
MAKE: Rica Lake TYPE: truck scale ACCURACY CLASS: III
MODEL: TQ 310 ID. NO: 91441421 CAPACITY: 100,000 lbs SECTIONS: 2
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: electronic PLATFORM: 10' X 12'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	4000	4000	0				
"	8000	8000	0	Left 1	20,000	20,000	0
"	12000	12000	0				
"	16000	16000	0				
Center	20000	20000	0	Left 2	20,000	20,000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 12205-1 and 2

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: [Signature] Date: June 9, 2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on June 9, 2005 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS